



## Intraoperative Image Guidance in Orbital and Lacrimal Surgery

*A Report by the American Academy of Ophthalmology*

Timothy J. McCulley, MD,<sup>1</sup> Vinay K. Aakalu, MD, MPH,<sup>2</sup> Jill A. Foster, MD,<sup>3,4</sup> Suzanne K. Freitag, MD,<sup>5</sup> Lora R. Dagi Glass, MD,<sup>6</sup> Seanna R. Grob, MD,<sup>7</sup> Jeremiah P. Tao, MD,<sup>8</sup> M. Reza Vagefi, MD,<sup>9</sup> Michael T. Yen, MD,<sup>10</sup> Michael K. Yoon, MD,<sup>11,12</sup> Stephen J. Kim, MD,<sup>13</sup> Edward J. Wladis, MD<sup>14</sup>

**Purpose:** To review the efficacy and safety of the use of intraoperative image guidance (IIG) in orbital and lacrimal surgery.

**Methods:** A literature search of the PubMed database was last conducted in November 2023 for English-language original research that assessed the use of any image guidance system in orbital and lacrimal surgery that included at least 5 patients. The search identified 524 articles; 94 were selected for full-text analysis by the panel. A total of 32 studies met inclusion criteria. The panel methodologist assigned a level II rating to 2 studies and a level III rating to 30 studies. No study met the criteria for level I evidence.

**Results:** Procedures reported on were as follows: fracture repair (n = 14), neoplasm and infiltrate biopsy or excision (n = 6), orbital decompression for Graves ophthalmopathy (n = 3), dacryocystorhinostomy (n = 1), and mixed etiology and procedures (n = 8). Four studies used more than one IIG system. One study that met level II evidence criteria compared the outcomes of orbital fracture repair with IIG (n = 29) and without IIG (n = 29). Borderline better outcomes were reported in the IIG group: 2% versus 10% with diplopia (P = 0.039) and 3% versus 10% with enophthalmos (P = 0.065). The other level II study compared the repair of fractures with navigation (n = 20) and without (n = 20). The group in which navigation was used had a measured mean volume reduction of 3.82 cm<sup>3</sup> compared with 3.33 cm<sup>3</sup> (P = 0.02), and there was a greater measured reduction in enophthalmos in the navigation group of 0.72 mm (P = 0.001). Although the remaining 30 assessed articles failed to meet level II criteria, all alleged a benefit from IIG. No complications were reported.

**Conclusions:** A small number of comparative studies suggest that there are improved outcomes when IIG is used in orbital fracture repair, but each study suffers from various limitations. No high-quality comparative studies exist for the management of lacrimal surgery, neoplastic disease, or decompression. Complications attributable to the use of IIG have not been identified, and IIG has not been analyzed for cost savings.

**Financial Disclosure(s):** Proprietary or commercial disclosure may be found after the references. *Ophthalmology* 2024;131:1333-1338 © 2024 by the American Academy of Ophthalmology



Supplemental material available at [www.aaojournal.org](http://www.aaojournal.org).

The American Academy of Ophthalmology prepares Ophthalmic Technology Assessments to evaluate new and existing procedures, drugs, and diagnostic and screening tests. The goal of an Ophthalmic Technology Assessment is to review systematically the available research for clinical efficacy, effectiveness, and safety. After review by members of the Ophthalmic Technology Assessment Committee, other Academy committees, relevant subspecialty societies, and legal counsel, assessments are submitted to the Academy's Board of Trustees for consideration as official Academy statements. The purpose of this assessment by the Ophthalmic Technology Assessment Committee Oculoplastics and Orbit

Panel is to review the literature on the efficacy and safety of the use of intraoperative image guidance (IIG) in orbital and lacrimal surgery.

### Background

Intraoperative image guidance refers to an ever-evolving imaging tool, most commonly computed tomography (CT) and magnetic resonance imaging (MRI), that is used intraoperatively to identify surgical targets, allowing for better identification of pathology and normal structures. It is used

in a growing number of surgical specialties, including oculoplastic surgery and overlapping specialties, neurosurgery, and otolaryngology. The purported benefits of IIG are precision of surgical dissection; accurate localization of lesions; and avoidance of critical vascular, neurological, and orbital structures.<sup>1–32</sup> Increased use of IIG has resulted in an increasing number of published reports assessing the benefits of IIG. These assessments are typically based primarily on surgeon impression without more objective, quantitative metrics or comparative control groups.

The evolving acumen of surgeons and the variety of technologies using different IIG systems add complexity to evaluations being made by different studies. For example, both CT and MRI data can be used for IIG. The means of image registration, the process of correlating digital images on a surgical navigation screen with the location on the patient, has varied and evolved substantially over time. Moreover, the variety of diseases and surgical techniques included in published reports makes data collation and comparison challenging. Although the evidence in the literature might seem plentiful and mostly positive, careful and meaningful interpretation of this technology is critical.

## Questions for Assessment

This assessment addressed the following questions: (1) Is there high-quality data that focuses on the safety and efficacy of IIG in orbital and lacrimal surgery and what conclusions can be drawn from it? (2) Does the benefit of IIG justify inherent costs?

## Description of Evidence

A literature search was last performed in November 2023 in the PubMed database for articles that assessed the use of any image guidance system in orbital and lacrimal surgery. The search terms for this assessment can be found in the [Appendix](#) (available at [www.aaojournal.org](http://www.aaojournal.org)). Articles were limited to original research published in the English language that included at least 5 patients. The search identified 524 articles, which were reviewed to select those that met the following inclusion criteria: (1) The study evaluated the safety or efficacy of adjunctive image guidance in orbital or lacrimal surgery, and (2) the study included a minimum of 1 month of follow-up. Ninety-four articles were deemed to be of sufficient clinical relevance and were selected for full-text review and abstraction by the panel. Articles were excluded if they included a cadaveric study, had an insufficient number of orbital cases, did not involve IIG, were assessments of technology as opposed to clinical outcomes, and did not contain original data (e.g., meta-analyses and review articles). Many articles were excluded on the basis of more than 1 criterion. A total of 32 studies met the study inclusion criteria.

The panel methodologist (V.K.A.) assessed the quality of the methodologies in each study and assigned each a level of evidence rating based on the scale developed by the 2011 Oxford Centre for Evidence-Based Medicine and adopted by the American Academy of Ophthalmology.<sup>33</sup> A level I

rating was assigned to well-designed and well-conducted randomized clinical trials. A level II rating was assigned to well-designed case-control and cohort studies and lower-quality randomized studies. A level III rating was assigned to case series, case reports, and lower-quality cohort and case-control studies. No study met the criteria for level I evidence. Two studies met the criteria for level II, and 30 studies met the criteria for level III.

## Published Results

There was a lack of consistency in IIG application and technology in the studies included in this assessment. The underlying pathologies addressed in the studies included in this assessment were fractures (14 articles), neoplasm and infiltrate biopsy or excision (6 articles), orbital decompression for Graves ophthalmopathy (3 articles), lacrimal surgery (1 article), and mixed etiology (8 articles).<sup>3</sup> The IIG manufacturer cited most often in the articles was Brainlab (n = 17), followed by Stryker (n = 5), Medtronic (n = 3), Boston Scientific Polaris (n = 1), Materialise (n = 1), and General Electric Medical Systems InstaTrak (n = 1). Four studies used more than one IIG system.

Orbital fracture repair was the most common IIG-assisted oculoplastic procedure reported in the published literature. In one of the level II studies, Cai et al<sup>1</sup> compared the outcomes of orbital fracture repair between 29 patients who underwent repair with the Kolibri surgical navigation device (Brainlab) and 29 patients who underwent repair without IIG. Better outcomes were reported in the Kolibri group: 2% versus 10% with diplopia ( $P = 0.039$ ) and 3% versus 10% with enophthalmos ( $P = 0.065$ ). The second level II study compared inferomedial fracture repair with navigation (n = 20) and without (n = 20).<sup>2</sup> Fracture size was inferred from the difference in orbit volume between the involved and uninvolved orbits and was similar between groups preoperatively. Patients were assigned to their group based on the site of surgery, one that was equipped with navigation and one that was not. The group for which navigation was used had a measured mean volume reduction of 3.82 cm<sup>3</sup> compared with 3.33 cm<sup>3</sup> ( $P = 0.02$ ) for the non-navigation group.<sup>2</sup> Moreover, there was a greater measured reduction in enophthalmos in the navigation group of 0.72 mm ( $P = 0.001$ ). Although the remaining 30 articles included in this assessment failed to meet level II criteria, all alleged benefit from IIG.<sup>3,5–32</sup> Apart from financial cost and time demands, no complications or other downsides were described.

In addition to the studies by Cai et al<sup>1</sup> and Parameswaran et al<sup>2</sup> detailed previously, other notable controlled studies assessed the use of IIG in orbital fracture repair. This includes 1 that compared outcomes of orbital fracture repair with and without the use of IIG and mirror image overlay.<sup>4</sup> There were 45 patients in both the study and control groups. Better outcomes were described for the patients in whom mirror image overlay was used and who had a more favorable revision rate (4% vs. 20%;  $P = 0.03$ ). He et al<sup>5</sup> compared outcomes in patients undergoing orbitozygomatic fractures repair using IIG and

custom implants (n = 11) with those not using IIG or custom implants (n = 39). They described outcomes as “perfect” in 100% of the IIG group compared with 74.3% in the traditional group. The groups were not similar with respect to the use of custom implants, nor was a statistical analysis provided.

Zong et al<sup>6</sup> compared fracture repair surgical duration and outcome with IIG (n = 40) and without (n = 30). By performing volumetric analysis of CT images before and after surgery, they measured better symmetry between the operative and fellow orbits with IIG (mean difference in orbital volume [0.57 ± 0.43 ml vs. 1.60 ± 0.78 ml; *P* = 0.022]). Interestingly, there was no difference in operative time, defined as the mean time from incision to closure (117.41 ± 36.74 minutes vs. 125.28 ± 40.73 minutes; *P* = 0.088). Note that this time did not include time required to set up the navigation system. Zavattero et al<sup>7</sup> compared patients who underwent fracture repair with IIG (n = 25) and without IIG (n = 30). Without any clearly defined outcome measures or a statistical comparison, they concluded that “significant orbital volume reduction in the reconstructed orbit could be achieved in the navigation group but not in the conventional group compared to the unaffected side.”

In a larger study by Scolozzi<sup>8</sup> of patients, most of whom underwent surgery with IIG for fractures or other trauma-related abnormality (n = 103), all patients were described as having “good and stable” outcomes.<sup>8</sup> Notably, the authors specifically stated that there were no complications, supporting the impression that there is limited clinical intraoperative risk to using IIG. Zhang et al<sup>9</sup> reported good results in 40 patients who underwent fracture repair using IIG. Likewise, Andrews et al<sup>10</sup> reported outcomes of 10 eyes in 8 patients who underwent orbital fracture repair using IIG. No complications were described, and all patients were noted to have normal “ocular function” postoperatively.<sup>10</sup> Baumann et al<sup>11</sup> also described 6 patients who achieved “symmetry” after unilateral fracture repair. There are several other largely anecdotal studies judging IIG to be beneficial in fracture repair.<sup>12–14</sup>

A number of researchers have reported on outcomes in patients undergoing orbital decompression with the assistance of IIG. Prevost et al<sup>15</sup> assessed patients who underwent orbital decompression either with IIG (205 eyes in 106 patients) or without IIG (145 eyes in 85 patients). They reported a statistically significant greater reduction in proptosis on the right (3.8 vs. 3.2 mm; *P* = 0.03) and a greater reduction on the left that was not statistically significant (3.6 vs. 3.4 mm; *P* = 0.54). They also reported a 40-minute increase in operative time when using IIG. Although the numbers are robust, the study lacks strength because it involved multiple surgeons and IIG was used based on the preference of the surgeon. Thus, the difference in outcome or surgical duration may relate to the surgeon and not to the use of IIG.

Dubin et al<sup>16</sup> compared outcomes in patients undergoing orbital decompression with IIG (n = 18) and without IIG (n = 27). Although the IIG group had a greater decrease

in proptosis (6.3 vs. 5.9 mm), this was not statistically significant. Tavassol et al<sup>17</sup> assessed the results of 12 patients (23 orbits) who underwent orbital decompression with IIG. Without precise outcome measures, they concluded that IIG was useful. Although not attributed specifically to IIG, 1 patient developed abducens nerve palsy. In a similar small study, Wu and Kahana<sup>18</sup> described good surgical results, with the desired amount of reduction in proptosis in 7 patients (11 orbits) who underwent decompression with IIG.

One application that is well accepted within the neurosurgical community that overlaps with orbital surgery is the adjunct use of IIG with the excision of skull base meningiomas, including those with extension to the orbit. The study by Maschke et al<sup>19</sup> was the single study identified in the literature search that specifically addressed this, and they described their experience with 30 patients. They assessed more than 1 imaging modality and asserted that bone surface–based registration was significantly more accurate than skin surface–based registration (mean 0.7 ± 0.4 mm and 1.9 ± 0.7 mm; *P* < 0.001). They also concluded that the extent of resection of the intraosseous component was significantly higher using CT plus MRI navigation compared with MRI alone (96% vs. 81%; *P* = 0.044). Others have reported their experiences using IIG for tumor location, excision, and reconstruction in the orbit.<sup>20–30</sup> Although published opinions in these studies were positive, a well-controlled study with clearly defined outcome measures in orbital neoplastic disease is lacking.

Only 1 study meeting the assessment criteria specifically addressed dacryocystorhinostomy. Reichel and Taxeidis<sup>31</sup> described using IIG in 17 eyes of 16 patients who underwent endoscopic dacryocystorhinostomy. Based on symptomatic improvement, all but 1 case were deemed successful. No complications attributable to IIG were described.

Study funding and author financial disclosures can be found in [Table S1](#) (available at [www.aaojournal.org](http://www.aaojournal.org)).

## Discussion

The use of IIG is commonplace in many surgical centers and is growing in popularity. In this assessment, the panel found more than 400 published reports that included descriptions of IIG in the orbit. Thirty met level III criteria, and only 2 met level II criteria. Inconsistency in the technology used and variability in surgical applications further dilute the utility of these studies. Moreover, the literature does not address whether anatomic familiarity and training influence clinical IIG outcomes for some surgeons more than others. Experienced surgeons who are more familiar with surgical anatomy and dissection might benefit less from IIG, if at all. One cadaveric study divided participants into novice (postgraduate year 2–4 residents; n = 6) and experienced (postgraduate year 5 residents and fellows; n = 4) cohorts.<sup>32</sup> Participants reconstructed iatrogenic cadaveric orbital fractures with or without the use of surgical navigation.

Novice participants had improved outcomes (more symmetric orbital volume) when using navigation without increased operative time, whereas more experienced participants did not have improved outcomes. This single study suggests that with surgeon experience the benefit of IIG diminishes.

## Conclusions

There is limited evidence supporting the adjunct use of IIG with various orbital and lacrimal procedures. There are few comparative studies that suggest improved outcomes when IIG is used in orbital fracture repair and decompression, but each study suffers from various limitations. Although the use of IIG is commonly described for the management of neoplastic disease, comparative studies are lacking. The lack of high-quality evidence in part relates to inconsistencies in the evolution of technology, the wide range of surgical applications, and differences in user experience and training. The few studies addressing surgical time indicate that although IIG neither extends nor shortens surgical time, equipment setup and removal extend facility time. Complications attributable to the use of IIG were not identified, and cost analyses are lacking.

## Footnotes and Disclosures

Originally received: May 10, 2024.

Final revision: May 13, 2024.

Accepted: May 13, 2024.

Available online: June 24, 2024. Manuscript no. OPHTHA-D-24-00950.

<sup>1</sup> Department of Ophthalmology, John P. McGovern Medical School, University of Texas Health Science Center, Houston, Texas.

<sup>2</sup> Department of Ophthalmology and Visual Sciences, University of Michigan, Ann Arbor, Michigan.

<sup>3</sup> Ophthalmic Surgeons and Consultants of Ohio, Columbus, Ohio.

<sup>4</sup> The Ohio State University, Ohio University Heritage College of Osteopathic Medicine, Athens, Ohio.

<sup>5</sup> Massachusetts Eye and Ear, Harvard Medical School, Boston, Massachusetts.

<sup>6</sup> Department of Ophthalmology, Edward S. Harkness Eye Institute, Columbia University Irving Medical Center, New York, New York.

<sup>7</sup> Department of Ophthalmology, University of California San Francisco, San Francisco, California.

<sup>8</sup> Gavin Herbert Eye Institute, University of California, Irvine School of Medicine, Irvine, California.

<sup>9</sup> Tufts University School of Medicine, Boston, Massachusetts.

<sup>10</sup> Cullen Eye Institute, Baylor College of Medicine, Houston, Texas.

<sup>11</sup> Department of Ophthalmology, Harvard Medical School, Boston, Massachusetts.

<sup>12</sup> Ophthalmic Plastic Surgery Service, Massachusetts Eye and Ear Infirmary, Boston, Massachusetts.

<sup>13</sup> Department of Ophthalmology, Vanderbilt University School of Medicine, Nashville, Tennessee.

<sup>14</sup> Ophthalmic Plastic Surgery, Lions Eye Institute, Department of Ophthalmology, Albany Medical Center, Albany (Slingerlands), New York.

## Future Research

There is a need for further research on the use of IIG in orbital and lacrimal surgery. Although the material costs of IIG technology are not defined in the literature, they are significant. Demands on personnel and facility time are impacted by equipment, which also comes with a substantial financial burden. Although the impact on actual operative time may be negligible, retrieval, assembly, and preparation need to be considered. Moreover, training of multiple staff, in addition to the surgeons themselves, is required if IIG is to be dependably available. However, IIG is commonplace in neurosurgery and otolaryngology, and in some institutions no additional equipment may be needed if it is used in orbital surgery. Also, if improved outcomes were to reduce the number of subsequent surgeries, there might be some cost savings. Future research should focus on better assessing clinical benefits and weighing this against the high-resource demand. This will require prospective randomization with quantifiable outcome measures. Specific diseases and procedures as well as surgeon characteristics need to be considered. Further investigation may identify the specific circumstances in which the use of IIG demonstrates improved clinical outcomes to the degree that justifies the substantial allocation of resources.

Disclosure(s):

All authors have completed and submitted the ICMJE disclosures form.

The author(s) have made the following disclosure(s):

T.J.M.: Leadership or fiduciary role in other board — American Society of Ophthalmic Plastic and Reconstructive Surgery, North American Neuro-Ophthalmology Society, American Academy of Ophthalmology, Association of University Professors of Ophthalmology.

V.K.A.: Grants — National Eye Institute, Department of Defense, Veterans Affairs; Consulting fees — EVOQ Therapeutics; Patents pending — Board of Trustees of University of Illinois, Regents of the University of Michigan; Stock or stock options — ViSo Therapeutics Inc., 8th Line III; Other financial or nonfinancial interests — Clinical Trial Site Investigator — Sling Therapeutics, Inc., F. Hoffman-La Roche AG, Amgen Inc.

S.K.F.: Royalties — Springer Publishing Company, Thieme Medical Publishers; Consulting fees — Horizon Therapeutics plc, Sling Therapeutics, Inc., Viridian Therapeutics, Inc., Immunovant, Inc., Medtronic, Inc., Merida/Third Rock Ventures, Poriferous, W.L. Gore & Associates, Inc., Kriya Therapeutics, Lassen Therapeutics, Janssen Pharmaceuticals.

L.R.D.G.: Royalties — Thieme Medical Publishers; Consulting fees — Ora Inc.; Advisory Board — Sling Therapeutics, Inc.; Leadership or fiduciary role — American Society of Ophthalmic Plastic and Reconstructive Surgery, Association of University Professors of Ophthalmology.

S.R.G.: Royalties — Springer Publishing Company; Support for attending meetings — Association of American Medical Colleges Women's Leadership in Medicine Seminar; Leadership or fiduciary role — Young American Society of Ophthalmic Plastic and Reconstructive Surgery Chair, Young American Society of Ophthalmic Plastic and Reconstructive Surgery Podcast workgroup; Receipt other services — Alcon Vision LLC event.

J.P.T.: Royalties — Springer Publishing Company; Patents pending — Devices and methods for surgical fire prevention.

M.R.V.: Leadership or fiduciary role — Chair of Ophthalmology, Tufts University School of Medicine, Oculoplastic Surgery Subspecialty



Representative, Association of University Professors of Ophthalmology-Fellowship Compliance Committee Board of Managers.

M.T.Y.: Grants — Viridian Therapeutics, Inc.; Consulting fees — Ipsen Innovation; Participation on a Data Safety Monitoring Board — Sling Therapeutics, Inc.

M.K.Y.: Participation on a Data Safety Monitoring Board — Versant Health; Leadership or fiduciary role in other board — Symposium director, American Society of Ophthalmic Plastic and Reconstructive Surgery.

Funded without commercial support by the American Academy of Ophthalmology.

Prepared by the Ophthalmic Technology Assessment Committee Oculoplastics and Orbit Panel and approved by the American Academy of Ophthalmology's Board of Trustees April 19, 2024.

HUMAN SUBJECTS: No human subjects were included in this study. All research adhered to the tenets of the Declaration of Helsinki. The requirement for informed consent was waived because of the retrospective nature of the study.

No animal subjects were used in this study.

Author Contributions:

Conception and design: McCulley, Aakalu, Foster, Freitag, Glass, Grob, Tao, Vagefi, Yen, Yoon, Kim, Wladis

Data collection: McCulley, Aakalu, Foster, Freitag, Glass, Grob, Tao, Vagefi, Yen, Yoon, Kim, Wladis

Analysis and interpretation: McCulley, Aakalu, Foster, Freitag, Glass, Grob, Tao, Vagefi, Yen, Yoon, Kim, Wladis

Obtained funding: N/A

Overall responsibility: McCulley, Aakalu, Foster, Freitag, Glass, Grob, Tao, Vagefi, Yen, Yoon, Kim, Wladis

Abbreviations and Acronyms:

**CT** = computed tomography; **IIG** = intraoperative image guidance;

**MRI** = magnetic resonance imaging.

Correspondence:


Andre Ambrus, MLIS, American Academy of Ophthalmology, Quality and Data Science, P.O. Box 7424, San Francisco, CA 94120-7424. E-mail: [aambrus@aao.org](mailto:aambrus@aao.org).

## References

- Cai EZ, Koh YP, Hing EC, et al. Computer-assisted navigational surgery improves outcomes in orbital reconstructive surgery. *J Craniofac Surg*. 2012;23:1567–1573.
- Parameswaran A, Pandey M, Panneerselvam E, et al. Does intraoperative navigation improve implant position accuracy in orbital fracture repair? *Facial Plast Surg Aesthet Med*. 2022.
- Campbell AA, Mahoney NR. Use of computer-assisted surgery in the orbit. *Orbit*. 2022;41:226–234.
- Bly RA, Chang SH, Cudejkova M, et al. Computer-guided orbital reconstruction to improve outcomes. *JAMA Facial Plast Surg*. 2013;15:113–120.
- He D, Li Z, Shi W, et al. Orbitozygomatic fractures with enophthalmos: analysis of 64 cases treated late. *J Oral Maxillofac Surg*. 2012;70:562–576.
- Zong CL, Shi YL, Jia JQ, et al. A retrospective study to compare the treatment outcomes with and without surgical navigation for fracture of the orbital wall. *Chin J Traumatol*. 2021;24:11–17.
- Zavattero E, Ramieri G, Rocchia F, Gerbino G. Comparison of the outcomes of complex orbital fracture repair with and without a surgical navigation system: a prospective cohort study with historical controls. *Plast Reconstr Surg*. 2017;139:957–965.
- Scolozzi P. Applications of 3D orbital computer-assisted surgery (CAS). *J Stomatol Oral Maxillofac Surg*. 2017;118:217–223.
- Zhang S, Gui H, Lin Y, et al. Navigation-guided correction of midfacial post-traumatic deformities (Shanghai experience with 40 cases). *J Oral Maxillofac Surg*. 2012;70:1426–1433.
- Andrews BT, Surek CC, Tanna N, Bradley JP. Utilization of computed tomography image-guided navigation in orbit fracture repair. *Laryngoscope*. 2013;123:1389–1393.
- Baumann A, Sinko K, Dorner G. Late reconstruction of the orbit with patient-specific implants using computer-aided planning and navigation. *J Oral Maxillofac Surg*. 2015;73:S101–S106.
- Chen CT, Pan CH, Chen CH, et al. Clinical outcomes for minimally invasive primary and secondary orbital reconstruction using an advanced synergistic combination of navigation and endoscopy. *J Plast Reconstr Aesthet Surg*. 2018;71:90–100.
- Chu YY, Yang JR, Pek CH, Liao HT. Application of real-time surgical navigation for zygomatic fracture reduction and fixation. *J Plast Reconstr Aesthet Surg*. 2022;75:424–432.
- Consorti G, Betti E, Catarzi L. Customized and navigated primary orbital fracture reconstruction: computerized operation neuronavigated surgery orbital recent trauma (CONSORT) protocol. *J Craniofac Surg*. 2022;33:1236–1240.
- Prevost A, Dekeister C, Caron P, et al. Outcomes of orbital decompression using surgical navigation in thyroid-associated ophthalmopathy. *Int J Oral Maxillofac Surg*. 2020;49:1279–1285.
- Dubin MR, Tabae A, Scruggs JT, et al. Image-guided endoscopic orbital decompression for Graves' orbitopathy. *Ann Otol Rhinol Laryngol*. 2008;117:177–185.
- Tavassol F, Kokemuller H, Muller-Tavassol C, et al. A quantitative approach to orbital decompression in Graves' disease using computer-assisted surgery: a compilation of different techniques and introduction of the "temporal cage". *J Oral Maxillofac Surg*. 2012;70:1152–1160.
- Wu CY, Kahana A. Stereotactic navigation with a registration mask in orbital decompression surgery: preliminary results. *Ophthalmic Plast Reconstr Surg*. 2015;31:440–444.
- Maschke S, Martinez-Moreno M, Micko A, et al. Challenging the osseous component of sphenoorbital meningiomas. *Acta Neurochir (Wien)*. 2019;161:2241–2251.
- Austin RE, Antonyshyn OM. Current applications of 3-D intraoperative navigation in craniomaxillofacial surgery: a retrospective clinical review. *Ann Plast Surg*. 2012;69:271–278.
- Bell RB, Markiewicz MR. Computer-assisted planning, stereolithographic modeling, and intraoperative navigation for complex orbital reconstruction: a descriptive study in a preliminary cohort. *J Oral Maxillofac Surg*. 2009;67:2559–2570.
- Camara JG, Nguyen LT, Fernandez-Suntay JP, et al. The use of a computer-assisted image-guided system (InstaTrak) in orbital surgery. *Ophthalmic Plast Reconstr Surg*. 2001;17:447–451.
- Enchev Y, Tzekov C, Ferdinandov D, et al. Neuronavigation in craniorbital neurosurgery - do we really need it? *Turk Neurosurg*. 2011;21:119–126.
- Markiewicz MR, Dierks EJ, Potter BE, Bell RB. Reliability of intraoperative navigation in restoring normal orbital dimensions. *J Oral Maxillofac Surg*. 2011;69:2833–2840.
- Nemec SF, Peloschek P, Schmook MT, et al. CT-MR image data fusion for computer-assisted navigated surgery of orbital tumors. *Eur J Radiol*. 2010;73:224–229.
- Unadkat SN, Rennie CE, Grant WE. Endoscopic transnasal image-guided approach to diagnosis in orbital apex and optic canal lesions. *J Laryngol Otol*. 2019;133:501–507.

27. Zhang WB, Mao C, Liu XJ, et al. Outcomes of orbital floor reconstruction after extensive maxillectomy using the computer-assisted fabricated individual titanium mesh technique. *J Oral Maxillofac Surg.* 2015;73:2065 e2061-2015.
28. Zhu M, Chai G, Lin L, et al. Effectiveness of a novel augmented reality-based navigation system in treatment of orbital hypertelorism. *Ann Plast Surg.* 2016;77:662–668.
29. Soh HY, Sun Q, Hu LH, et al. Accuracy of globe-sparing orbital reconstruction using individually bent titanium mesh: a comparative study. *J Plast Reconstr Aesthet Surg.* 2022;75: 1971–1978.
30. Tel A, Murta F, Sembronio S, et al. Virtual planning and navigation for targeted excision of intraorbital space-occupying lesions: proposal of a computer-guided protocol. *Int J Oral Maxillofac Surg.* 2022;51:269–278.
31. Reichel O, Taxeidis M. Use of an image-guided navigation system for routine endonasal endoscopic dacryocystorhinostomy. *J Laryngol Otol.* 2019;133:685–690.
32. Davis KS, Vosler PS, Yu J, Wang EW. Intraoperative image guidance improves outcomes in complex orbital reconstruction by novice surgeons. *J Oral Maxillofac Surg.* 2016;74: 1410–1415.
33. Oxford Centre for Evidence-Based Medicine Levels of Evidence Working Group, The Oxford 2011 Levels of Evidence. Oxford Centre for Evidence-Based Medicine. <https://www.Cebm.Net/wp-content/uploads/2014/06/cebm-levels-of-evidence-2.1.Pdf>. Accessed April 19, 2024.

## Pictures & Perspectives



### Squamous Cell Papilloma of the Conjunctiva

A 28-year-old woman presented to the ophthalmology department with a 3-year history of a tumor in her left eye. Examination revealed an isolated pink tumor in the conjunctiva with an associated large vessel (A). OCT angiography showed blood vessels in the shape of “coral” (B). After complete resection, pathological examination showed conjunctival papilloma (C). Conjunctival papilloma is an acquired benign tumor that originates from the conjunctival stratified squamous epithelium and usually progresses slowly. Most patients require surgical resection. Lesions may also be treated pharmacologically or via cryotherapy. (Magnified version of Figure A-C is available online at [www.aaojournal.org](http://www.aaojournal.org)).

XIULAN ZHAO, MD<sup>1</sup>  
XUAN MENG, MB<sup>1</sup>  
GUOPING WANG, MD<sup>2</sup>

<sup>1</sup>Ineye Hospital of Chengdu University of Traditional Chinese Medicine, Chengdu, Sichuan, China; <sup>2</sup>Eye School of Chengdu University of Traditional Chinese Medicine, Chengdu, Sichuan, China